



## Bike ergometer for pediatric use – ergoselect 150

- special saddle adjustment for children
- length adjustable cranks (12 – 17 cm)
- clamping fixture allows easy change of saddle – for use with adults
- digital and analog interface to ECG and PC-ECG
- automatic, non-invasive bloodpressure measurement (option, upgradable)



bike ergometer • supine ergometer • stressecho-supine ergometer • cardiac rehabilitation

## Technical Data

<b>Ergometer</b>	<b>ergoselect 150</b>
Brake system	microprocessor controlled eddy current brake
Load	6 – 999 Watt, speed independent
Accuracy	according to DIN VDE 0750-238
Speed range	30 – 130 U/min
Saddle height adjustment	continuous, mechanic
Handlebar adjustment	inclination: 360° / height: 90 – 126 cm
Patient weight (max.)	max. 60 kg (pediatric saddle) / max. 160 kg (adult saddle)
<b>Control unit</b>	
Display / patientdisplay	load, rpm, speed, time, blood pressure, SpO2, heart rate (LCD) / rpm (LED)
Keyboard	foil keyboard
Grafic display (load, heartrate)	○
<b>Stresstest programs</b>	
User definable	10
Predefined programs (WHO, Hollmann etc.)	5
Manual load adjustment	●
<b>Programs for training</b>	
Puls steady state training (integr. POLAR-receiver)	○
Predefined fitness tests	○
<b>Options</b>	
Automatic bloodpressure measurement	○
Oxygen saturation measurement	○
Standard saddle with clamp fixture	○
<b>Interface</b>	
Digital (RS-232, USB) / analog (load) / remote start	● / ○ / ○
Connection to ERS Ergoline Rehabilitation System	○
<b>Miscellaneous</b>	
Dimensions, max. (L x W x H)	ca. 90 cm x 46 cm x 133 cm (max.)
Weight	55 kg
Power supply	100 – 240 V / 50 – 60 Hz / 50 VA max.

● standard ○ option

# ergoline

ergoline GmbH  
Lindenstrasse 5  
72475 Bitz  
Germany

Tel.: +49-(0)-7431 - 9894 - 0  
Fax: +49-(0)-7431 - 9894 - 128  
email: info@ergoline.com  
internet: www.ergoline.com

From development to production, all ergoline products are subject to a quality management system certified to ISO 9001:2008 and EN ISO 13485:2012. All products are CE-marked and fulfill the requirements of the medical device directive 93/42/EEC.

The data provided is based on the information valid at the time of printing.  
Subject to modifications.